



HERMLE Labortechnik GmbH Centrifuges and Rotor Compatibility

TubeSpin® Bioreactor 450

The following rotors of the centrifuge from Hermle are compatible for the TubeSpin® Bioreactor 450.



Underbench



Floor-Standing

ZK 496



221.65 V20



621.000



616.102



Swing-Out Rotor		221.65 V20	
Bucket	Bucket Lid	Adapter	Capacity
621.000	616.102	220.70.133	4 x 87450



Z 446



Z 446K



221.36 V20



616.200



616.102

Swing-Out Rotor		221.36 V20	
Bucket	Bucket Lid	Adapter	Capacity
616.200	616.102	220.70.133	4 x 87450

For more technical information on centrifuges, rotors and adapters from
 HERMLE Labortechnik GmbH
 please visit www.herml-labortechnik.de and/or contact info@hermlt.de.



Technical Data
TubeSpin® Bioreactor 450 / 600

Measurements	87450	87600
Volume graduation mL	450	600
Length mm	145	183
Diameter mm	96	100
Max. RCF x g	3'500	3'500
Form	conical	conical
Optimal filling volume mL	>250	>400
Shaker: Recom. orbit/shaking throw mm	50	50
Shaker: Recom. speed rpm	150	150
Screw cap „solid“	87355	87355
Screw cap „filter“	87356	87356

Disclaimer

TPP products are for research use only and not for clinical, diagnostic or therapeutic use. All products are intended for use by trained personnel that are familiar with safe laboratory practices.

TPP assumes no responsibility for damage or defects resulting from improper or unauthorized use. It is the responsibility of the user to store, handle, and use the products in accordance with the instructions provided.

TPP does not warrant the completeness or accuracy of this TechDoc. TPP's recommendations are intended as general guidelines and may not cover all possible scenarios. TPP shall not be liable for any indirect, incidental, consequential, or special damages arising out of the use or inability to use the information in this TechDoc.

Swiss law governs these terms of use and any resulting legal matters.